



Verify IGRT Plans and Avoid Gated Treatment Errors



SIMULATE PATIENT BREATHING

The QUASAR™ Respiratory Motion Phantom is a programmable state-of-the-art breathing simulator for conducting quality assurance testing on radiotherapy systems. Using a personal computer to communicate with the QUASAR™ Respiratory Motion Phantom, the software application downloads patient respiratory waveforms to the phantom, simulating the breathing function.

KEY FEATURES


- ▶ Operate the phantom directly from a PC or laptop
- ▶ Stream patient-specific respiratory motion waveforms
- ▶ Interchangeable inserts for multiple testing applications
- ▶ Works with a number of motion tracking systems
- ▶ Includes an unlimited multiple site software license

BENEFITS

- ▶ Communicate with phantom through local area network
- ▶ Requires no additional programming or customization
- ▶ Ability to rotate inserts as they translate to create 3D motion
- ▶ Compatible with VXP, .CSV, .TXT, .DCM, .LOG, .DAF, .IMA
- ▶ Install the software on an unlimited number of computers

Used for regularly scheduled testing, as well as commissioning new systems and upgrades, and testing repairs, the Respiratory Motion Phantom is designed to move cylindrical inserts in the superior/inferior direction within a body shaped oval with both varying speed and amplitude. Several different moving inserts are available for a variety of applications. Insert motion is linked to a moving Chest Wall Platform designed to carry a respiratory tracking device.

The QUASAR™ Respiratory Motion Phantom is designed to work with different therapy delivery systems and is compatible with kV and MV CT imagers.



Above: Phantom Control Module

PATIENT-SPECIFIC SOFTWARE

Included with the phantom is the QUASAR™ Respiratory Motion QA Software application which allows you to import, create, edit and save respiratory waveforms. Easily import patient-specific waveforms from a number of respiratory gating machines including Varian Real-time Position Management™ (RPM), Anzai, Cyberknife, Philips, Respisens, and Siemens. Quickly create custom waveforms or import unique waveforms produced using tab delimited spreadsheet files.



Accuracy. Confidence.™

Conveniently edit waveforms using a wide range of functions include adjusting the amplitude, stretching or compressing the timeline and filtering out high frequency noise, low frequency drift and cardiac signals.

The software is compatible with Windows 7 or better.

MINIMUM TECHNICAL REQUIREMENTS


- ▶ **Operating System:** Windows 7 or higher
- ▶ **Processor:** Intel Core™ i7 or better
- ▶ **Hard Disk Space:** 5 GB minimum (64-bit)
- ▶ **RAM:** 1 gigabyte (GB) or more
- ▶ **Screen Resolution:** 1280 x 1024 pixels or higher
- ▶ **Graphics Card:** Intel HD Integrated Graphics or better

SPECIFICATIONS

- ▶ 7 kg Acrylic Body Oval: 30 cm W x 20 cm H x 12 cm L
- ▶ 5 kg Drive Unit: 20 cm L x 15 cm W x 12 cm H
- ▶ Total weight is approximately 20 kg with all options
- ▶ 2 openings in body oval: 8 cm diameter each, for drive unit and moving insert
- ▶ Includes cedar insert: drilled for 2 cm diameter acrylic ion chamber holder
- ▶ Chest wall platform: 13 cm diameter, carries up to 500 g
- ▶ Power supply: Input, 100 – 240 V AC, 47 – 63 Hz
- ▶ Output, 24 V DC 2.1 A, 50 W.
- ▶ Approvals: CE, UL/CSA 60950-1



Above: Optional Rotation Stage accessory for adding 3D motion



Above: Wave Editor Module

ORDERING INFORMATION

- | | |
|----------|------------------------------------|
| 100-1011 | QUASAR™ Respiratory Motion Phantom |
| | ▶ 1-Software License |
| | ▶ 1-Phantom |
| | ▶ User's guide |

OPTIONAL ACCESSORIES

- | | |
|----------|---------------------------------------|
| 500-3330 | Rotation Stage for 3D motion |
| 500-2003 | Phantom Heavy-Duty Shipping Case 8 kg |

OPTIONAL INSERTS

- | | |
|----------|---|
| 500-3305 | Acrylic Insert |
| 500-3311 | Cedar Insert with Solid Tumour |
| 500-3312 | Cedar Insert with Solid Tumour – Drilled |
| 500-3313 | Cedar Lung Tumour Insert (Split) |
| 500-3314 | Hollow Insert with Screw Cap |
| 500-3315 | Film Cassette Insert |
| 500-3317 | 4D CT Imaging Insert |
| 500-3318 | PET/CT Insert |
| 500-3322 | MP Insert Adapter |
| 500-3331 | Offset Cedar Insert with Solid Tumour |
| 500-3332 | Offset Cedar Insert with Solid Tumour - Drilled |
| 500-3333 | Offset Cedar Lung Tumour Insert (Split) |
| 500-3395 | Cedar Ion Chamber Holder |

© 2017 Modus Medical Devices Inc. All Rights Reserved. Specifications subject to change without notice. Modus QA is not responsible for errors or omissions. This product manufactured under license from CIRS, Inc., Norfolk, Virginia, USA, Pat. 7151253. PDS#100-1011, REV#10/26

Modus QA

1570 North Routledge Park,
London, Ontario Canada N6H 5L6

[\[e\] info@modusQA.com](mailto:info@modusQA.com)

[\[w\] www.modusQA.com](http://www.modusQA.com)

Toll Free: +1 (866) 862-9682 (North America)

Phone: +1 (519) 438-2409

Fax: +1 (519) 643-0127

GET SOCIAL



modusQA.com